

SELF-EXPANDING STENT GRAFTS

Company Name	Product Name	US FDA Indicated Use	Materials Used	Introducer Size (F)	Endhole (inch)	Stent Diameter (mm)	Stent Length (mm)	Delivery System Length (cm)
Bard Peripheral Vascular, Inc.	Flair Endovascular Stent Graft	Stenosis at venous anastomosis of ePTFE or other synthetic AV grafts	Nitinol/ePTFE	9	0.035	6, 7, 8, 9	30, 40, 50	80
Gore & Associates	Gore Viabahn Endoprosthesis, Gore Viabahn Endoprosthesis With Heparin Bioactive Surface	5–8 indicated for SFA; 5–13 indicated for iliac artery	Nitinol/ePTFE	6–12	0.014/0.018 or 0.035	5, 6, 7, 8, 9, 10, 11, 13	25, 50, 100, 150	75, 120
	Gore Viabil Biliary Endoprosthesis	Malignant biliary strictures ^a		10 percutaneous, 8.5 endoscopic	0.035	8, 10	40, 60, 80, 100	40 percutaneous, 200 endoscopic
	Gore Viatorr TIPS Endoprosthesis	TIPS, tracheobronchial		10	0.038	8, 10, 12	60, 70, 80, 90, 100	75

AV, arteriovenous; ePTFE, expanded polytetrafluoroethylene; SFA, superficial femoral artery; TIPS, transjugular intrahepatic portosystemic shunt.

^aThe Gore Viabil Biliary Endoprosthesis is contraindicated for all cardiovascular applications.